

Beamline Hazard Analysis - Beamline X26A

LS-BHA-X26A

Revision: 02

Effective: 10/18/11

Page 1 of 1

The following tasks are included in the beamline safety review. Any operations not listed here, or performed by others than those listed, need additional review by ES&H personnel and/or beamline personnel.

Prepared By: A. Lanzirotti/S. Wirick

Approved By: Lori Stiegler

* Approval signatures on file with master copy

Tasks / Equipment	Performed by:		Controls
	Beamline Staff	Users	
Inert compressed gas	Y	N	<ul style="list-style-type: none"> Compressed gas training – TQ-COMPGAS1 or NSLS User training Transport cylinders using a cylinder cart Properly secure cylinders Wear safety glasses with side shields while connecting lines Leak check gas lines
Using the LN2 Fill Station or pressurized transfer	Y	N	<ul style="list-style-type: none"> LS-PROC-LN2FILL-MANL training for fill station Use full face shield along with goggles or safety glasses Long pants and sleeves - avoid clothing which can trap cryogenics (pockets, cuffs) Cryogen or heavy gloves Enclosed shoes
Cryogenic autofill system	Y	Y	<ul style="list-style-type: none"> Ensure hutch ventilation system is operating
Beryllium use	Y	Y	<ul style="list-style-type: none"> Handle articles with gloves, Leave area and contact OpCo if damaged
Cutting/razor blades	Y	Y	<ul style="list-style-type: none"> Be aware of hand positioning Use safety knife when possible Consider using cut resistant gloves for frequent cutting tasks